


		IEC 118-7: 1994 (2cc)	IEC 118-0: 1994 (711 OES)	ANSI S3.22-1987 (2cc)	Limits
Saturation	Peak	123	128	123	±3
OSPL90	F Reference	114	124	123	±3
	HF Average			118	±3
P Control at RTF			-14	-10	±2
Full-on Gain	Peak	46	56	46	±4
	F Reference	42	52	45	±4
	HF Average			43	±4
Nominal Reference Test Gain (RTG)		35	45	41	
Frequency Response				200 - 5000	
H Control at:	300 Hz		-15		±2
	500 Hz			-12	±2
Volume Control Range			> 40		
Total Harmonic Distortion at RTG:					
70 dB SPL In	500 Hz				≤ 5
	800 Hz				≤ 5
	1600 Hz	1	1	1	≤ 5
Equivalent noise level		22	24	22	≤ 25
Induction Coil Sensitivity					
FOG at 10mA/m	dB SPL	95		101	±4
RTG at 31.6mA/m	dB SPL		114		±4
Supply Current at RTF	input dB SPL	60		65	≤ 1.4
	mA	1.2		1.2	
Battery Life (Typical)		Type 675 Zinc Air		430	430
Reference Test Frequency		1600	1600	1000	



## features:

This is a small, medium powered, behind-the-ear, hearing instrument with two pre-set trimmer controls:

A Low Frequency Cut Control 'H' and a Peak Clip Control 'P'.

The normal application is for a patient with a mild to moderate hearing loss.

### Separate switch 3 positions:

O - off

T - telecoil

M - microphone

### Battery Type 675

IEC PR44 (675) Zinc-Air

- Screw Elbow (Hook)
- Technician operated Pre-Set controls, P and H.
- Available in beige, brown and grey colours.
- Forward facing microphone, situated above the output nozzle.

### Acoustic Elbow (Hook)

We reserve the right to alter specifications of hearing instruments without advance notice.

